

# **ACCREDITATION SCHEME FOR LABORATORIES**

# **Technical Notes NDT 002**

Guidelines for the Recognition of Non-Destructive Testing (NDT) Personnel Certification for SAC-SINGLAS Assessments

#### 1. Introduction

This document provides guidelines for the recognition of personnel certification of staff for non-destructive testing (NDT) laboratories accredited under SAC-SINGLAS.

## 2. NDT Personnel Certification requirements for approved signatories

- 2.1 The NDT personnel certification levels shall be those as defined in the various national/international NDT personnel certification schemes.
- 2.2 SAC-SINGLAS approved signatories for Non-destructive Testing except for Positive Material Identification (PMI) and X-ray Diffraction (XRD), shall be certified by any of the following bodies and have 2 years of relevant supervisory experience:
  - a) American Society for Non-destructive Testing (ASNT Level III)
  - b) ASNT Central Certification Programme (ACCP Level II & III)
  - c) Australia Institute of Non-destructive Testing (AINDT Level II & III)
  - d) British Institute for Non-destructive Testing (BINDT) (PCN Level II & III)
  - e) TWI Certification Ltd (CSWIP Level II & Level III)
  - f) NDT Society Singapore (NDT SS) (SGNDT Level II & Level III)
- 2.3 SAC-SINGLAS approved signatories for the PMI and XRD shall meet the requirement of SAC-SINGLAS 001 with 2 years of relevant supervisory experience and evidence of training by Original Equipment Manufacturer (OEM) or its authorized agent on the operation of the equipment.
- 2.4 The above list of recognised NDT personnel certification schemes shall be approved and reviewed as required by the SAC-SINGLAS Technical Committee for Physical Engineering.

### 3. NDT Personnel Certification requirements for testing operators

- 3.1 In-house (employer based Level II and above) or recognised NDT personnel certification as per clause 2.2.
- 3.2 NDT personnel who carry out Phased Array Ultrasonic Testing and/or Time of Flight Diffraction Method(s) shall be certified as per clause 2.2 for the specific method.